


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121622\
 Data File : VX033449.D
 Acq On : 16 Dec 2022 08:26
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X121222W.M
 Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 Last Update : Mon Dec 12 12:37:35 2022



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1633

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.5	3777	PASS
75	95	30	60	52.8	10260	PASS
95	95	100	100	100.0	19414	PASS
96	95	5	9	6.9	1332	PASS
173	174	0.00	2	1.1	176	PASS
174	95	50	100	82.1	15941	PASS
175	174	5	9	6.8	1085	PASS
176	174	95	101	95.0	15151	PASS
177	176	5	9	6.7	1020	PASS