

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121623\
 Data File : VX039265.D
 Acq On : 16 Dec 2023 05:26
 Operator : JC/MD
 Sample : 05845-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 YCHE6

Quant Time: Dec 16 20:15:37 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML120523WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Dec 12 03:50:51 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	240585	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	222572	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	103724	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	59273	31.411	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	62.820%
7) Chloroethane-d5	1.672	69	48868	35.036	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	70.080%
11) 1,1-Dichloroethene-d2	2.307	65	32067	38.713	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	77.420%
21) 2-Butanone-d5	4.452	46	146637	85.847	ug/L	0.01
Spiked Amount	100.000	Range	40 - 130	Recovery	=	85.850%
24) Chloroform-d	5.056	84	142107	42.541	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.080%
26) 1,2-Dichloroethane-d4	5.958	65	98791	45.665	ug/L	0.01
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.340%
32) Benzene-d6	5.970	84	280276	40.803	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	81.600%
36) 1,2-Dichloropropane-d6	7.306	67	91941	41.591	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	83.180%
41) Toluene-d8	8.647	98	250908	40.307	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	80.620%
43) trans-1,3-Dichloroprop...	8.952	79	41568	36.951	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	73.900%
47) 2-Hexanone-d5	9.378	63	104677	79.282	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	79.280%
56) 1,1,2,2-Tetrachloroeth...	11.189	84	133681	42.629	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	85.260%
66) 1,2-Dichlorobenzene-d4	12.317	152	87066	44.408	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.820%
Target Compounds						
5) Vinyl chloride	1.374	62	8638	4.370	ug/L	85
12) 1,1-Dichloroethene	2.319	96	145064	93.597	ug/L #	79
13) Acetone	2.374	43	68942	42.014	ug/L	95
19) 1,1-Dichloroethane	3.611	63	45844	14.198	ug/L	99
20) cis-1,2-Dichloroethene	4.489	96	28028	14.797	ug/L	93
22) 2-Butanone	4.550	43	73951	36.721	ug/L	99
30) 1,1,1-Trichloroethane	5.385	97	10585	3.920	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121623\
Data File : VX039265.D
Acq On : 16 Dec 2023 05:26
Operator : JC/MD
Sample : 05845-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
YCHE6

Quant Time: Dec 16 20:15:37 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML120523WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Dec 12 03:50:51 2023
Response via : Initial Calibration

