

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121623\
 Data File : VX039284.D
 Acq On : 16 Dec 2023 12:48
 Operator : JC/MD
 Sample : 05845-02DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 55 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 YCHD8DL

Quant Time: Dec 16 21:11:56 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML120523WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Dec 12 03:50:51 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.757	114	202796	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	188575	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	93745	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.367	65	70316	44.207	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	88.420%
7) Chloroethane-d5	1.648	69	32117	27.317	ug/L	-0.02
Spiked Amount	50.000	Range	70 - 130	Recovery	=	54.640%#
11) 1,1-Dichloroethene-d2	2.300	65	34440	49.325	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	98.660%
21) 2-Butanone-d5	4.465	46	137352	95.395	ug/L	0.02
Spiked Amount	100.000	Range	40 - 130	Recovery	=	95.390%
24) Chloroform-d	5.056	84	134788	47.869	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.740%
26) 1,2-Dichloroethane-d4	5.952	65	96318	52.818	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.640%
32) Benzene-d6	5.970	84	281709	48.405	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.820%
36) 1,2-Dichloropropane-d6	7.305	67	90292	48.209	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.420%
41) Toluene-d8	8.647	98	249539	47.314	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.620%
43) trans-1,3-Dichloroprop...	8.951	79	40834	42.843	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	85.680%
47) 2-Hexanone-d5	9.384	63	102549	91.673	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	91.670%
56) 1,1,2,2-Tetrachloroeth...	11.189	84	128202	48.252	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	96.500%
66) 1,2-Dichlorobenzene-d4	12.317	152	87913	49.614	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.220%
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.312	96	67631	51.768	ug/L	88
13) Acetone	2.386	43	39991	28.912	ug/L	95
19) 1,1-Dichloroethane	3.605	63	11101	4.079	ug/L	96
27) 1,2-Dichloroethane	6.092	62	4738	2.211	ug/L	98
30) 1,1,1-Trichloroethane	5.385	97	15882	6.941	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121623\
Data File : VX039284.D
Acq On : 16 Dec 2023 12:48
Operator : JC/MD
Sample : 05845-02DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 55 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
YCHD8DL

Quant Time: Dec 16 21:11:56 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM120523WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Dec 12 03:50:51 2023
Response via : Initial Calibration

