

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121623\
 Data File : VX039294.D
 Acq On : 16 Dec 2023 16:39
 Operator : JC/MD
 Sample : 05845-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 65 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 YCHE7

Quant Time: Dec 16 21:13:21 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML120523WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Dec 12 03:50:51 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	217955	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	202369	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	96181	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	69005	40.365	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	80.740%
7) Chloroethane-d5	1.666	69	52494	41.544	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	83.080%
11) 1,1-Dichloroethene-d2	2.306	65	34088	45.425	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	90.860%
21) 2-Butanone-d5	4.452	46	147465	95.295	ug/L	0.01
Spiked Amount	100.000	Range	40 - 130	Recovery	=	95.300%
24) Chloroform-d	5.056	84	142890	47.217	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.440%
26) 1,2-Dichloroethane-d4	5.958	65	98240	50.125	ug/L	0.01
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.260%
32) Benzene-d6	5.976	84	292171	46.781	ug/L	0.01
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.560%
36) 1,2-Dichloropropane-d6	7.305	67	93549	46.543	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	93.080%
41) Toluene-d8	8.647	98	260100	45.955	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.920%
43) trans-1,3-Dichloroprop...	8.952	79	41435	40.510	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	81.020%
47) 2-Hexanone-d5	9.384	63	105403	87.802	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	87.800%
56) 1,1,2,2-Tetrachloroeth...	11.189	84	130709	45.842	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	91.680%
66) 1,2-Dichlorobenzene-d4	12.317	152	87218	47.975	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.940%
Target Compounds						
					Qvalue	
5) Vinyl chloride	1.374	62	5975	3.337	ug/L	# 82
12) 1,1-Dichloroethene	2.319	96	78090	55.616	ug/L	92
13) Acetone	2.380	43	58651	39.453	ug/L	97
19) 1,1-Dichloroethane	3.611	63	37646	12.870	ug/L	93
20) cis-1,2-Dichloroethene	4.489	96	24432	14.238	ug/L	97
22) 2-Butanone	4.556	43	76962	42.184	ug/L	99
30) 1,1,1-Trichloroethane	5.385	97	5761	2.346	ug/L	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121623\
Data File : VX039294.D
Acq On : 16 Dec 2023 16:39
Operator : JC/MD
Sample : 05845-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 65 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
YCHE7

Quant Time: Dec 16 21:13:21 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM120523WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Dec 12 03:50:51 2023
Response via : Initial Calibration

