

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121624\
 Data File : VX044321.D
 Acq On : 16 Dec 2024 14:11
 Operator : JC/MD
 Sample : P5261-20
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP-02-20241210

Quant Time: Dec 17 00:50:03 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 12 04:28:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	114448	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	220754	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	195441	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	77156	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	99078	49.375	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	98.740%		
35) Dibromofluoromethane	5.385	113	71639	46.856	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	93.720%		
50) Toluene-d8	8.646	98	268405	52.035	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	104.080%		
62) 4-Bromofluorobenzene	11.079	95	89290	49.540	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	99.080%		
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.318	101	1191	0.853	ug/l #	87
16) Acetone	2.379	43	860	1.051	ug/l	96
20) Methylene Chloride	2.788	84	410	0.267	ug/l #	55
44) Trichloroethene	7.122	130	16942	11.322	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121624\
Data File : VX044321.D
Acq On : 16 Dec 2024 14:11
Operator : JC/MD
Sample : P5261-20
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP-02-20241210

Quant Time: Dec 17 00:50:03 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 12 04:28:14 2024
Response via : Initial Calibration

