

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121624\
 Data File : VX044325.D
 Acq On : 16 Dec 2024 15:43
 Operator : JC/MD
 Sample : P5261-15DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE125D2-20241210DL

Quant Time: Dec 17 00:51:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 12 04:28:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	127800	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	235549	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	201274	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	78713	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	106337	47.456	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	94.920%		
35) Dibromofluoromethane	5.379	113	77095	47.257	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	94.520%		
50) Toluene-d8	8.647	98	286322	52.022	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	104.040%		
62) 4-Bromofluorobenzene	11.079	95	89801	46.694	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	93.380%		
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.325	101	1060	0.680	ug/l	94
12) 1,1-Dichloroethene	2.307	96	405	0.283	ug/l	83
16) Acetone	2.386	43	721	0.789	ug/l	100
27) cis-1,2-Dichloroethene	4.489	96	743	0.384	ug/l	85
44) Trichloroethene	7.123	130	55846	34.977	ug/l	94
64) Tetrachloroethene	9.275	164	1655	1.239	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121624\
Data File : VX044325.D
Acq On : 16 Dec 2024 15:43
Operator : JC/MD
Sample : P5261-15DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE125D2-20241210DL

Quant Time: Dec 17 00:51:39 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 12 04:28:14 2024
Response via : Initial Calibration

