

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121624\
 Data File : VX044326.D
 Acq On : 16 Dec 2024 16:06
 Operator : JC/MD
 Sample : P5261-18DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE120D2-20241210DL

Quant Time: Dec 17 00:52:07 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 12 04:28:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	118530	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	215461	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	191459	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	73600	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	96019	46.203	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	92.400%
35) Dibromofluoromethane	5.385	113	69622	46.655	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.320%
50) Toluene-d8	8.647	98	264552	52.548	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	105.100%
62) 4-Bromofluorobenzene	11.079	95	82769	47.050	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	94.100%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.319	101	14623	10.110	ug/l	98
27) cis-1,2-Dichloroethene	4.507	96	513	0.286	ug/l #	52
44) Trichloroethene	7.123	130	55411	37.940	ug/l	94
64) Tetrachloroethene	9.275	164	1214	0.955	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121624\
Data File : VX044326.D
Acq On : 16 Dec 2024 16:06
Operator : JC/MD
Sample : P5261-18DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE120D2-20241210DL

Quant Time: Dec 17 00:52:07 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 12 04:28:14 2024
Response via : Initial Calibration

