

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121624\
 Data File : VX044332.D
 Acq On : 16 Dec 2024 18:26
 Operator : JC/MD
 Sample : P5280-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BPOW6-7-20241212

Quant Time: Dec 17 00:54:52 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 12 04:28:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	113355	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	210532	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	182215	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	71993	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	92471	46.527	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	93.060%		
35) Dibromofluoromethane	5.379	113	68856	47.222	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	94.440%		
50) Toluene-d8	8.647	98	255542	51.947	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	103.900%		
62) 4-Bromofluorobenzene	11.079	95	81479	47.401	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	94.800%		
Target Compounds						
16) Acetone	2.392	43	804	0.992	ug/l	93
44) Trichloroethene	7.122	130	5438	3.811	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121624\
Data File : VX044332.D
Acq On : 16 Dec 2024 18:26
Operator : JC/MD
Sample : P5280-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BPOW6-7-20241212

Quant Time: Dec 17 00:54:52 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 12 04:28:14 2024
Response via : Initial Calibration

