

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121625\
 Data File : VX048895.D
 Acq On : 16 Dec 2025 18:29
 Operator : JC/MD
 Sample : VX1216WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1216WBL02

Quant Time: Dec 17 03:27:02 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 17 03:22:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	77608	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	166359	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	179509	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	90797	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	75414	58.134	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	116.260%
35) Dibromofluoromethane	5.367	113	61771	49.956	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.920%
50) Toluene-d8	8.635	98	217522	49.274	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.540%
62) 4-Bromofluorobenzene	11.061	95	102248	64.513	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	129.020%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121625\
Data File : VX048895.D
Acq On : 16 Dec 2025 18:29
Operator : JC/MD
Sample : VX1216WBL02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1216WBL02

Quant Time: Dec 17 03:27:02 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121625W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 17 03:22:16 2025
Response via : Initial Calibration

