

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121625\
 Data File : VX048900.D
 Acq On : 16 Dec 2025 20:13
 Operator : JC/MD
 Sample : Q3776-02RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-20251203RE

Quant Time: Dec 17 03:29:29 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 17 03:22:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	128283	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	222478	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	196046	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	81284	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	91118	42.493	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	84.980%
35) Dibromofluoromethane	5.373	113	70964	42.914	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	85.820%
50) Toluene-d8	8.635	98	252928	42.842	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	85.680%#
62) 4-Bromofluorobenzene	11.061	95	85870	40.513	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	81.020%#
Target Compounds						
17) Carbon Disulfide	2.526	76	1706	0.422	ug/l #	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121625\
Data File : VX048900.D
Acq On : 16 Dec 2025 20:13
Operator : JC/MD
Sample : Q3776-02RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-20251203RE

Quant Time: Dec 17 03:29:29 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121625W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 17 03:22:16 2025
Response via : Initial Calibration

