

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121625\
 Data File : VX048903.D
 Acq On : 16 Dec 2025 21:16
 Operator : JC/MD
 Sample : Q3850-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Dec 17 03:30:25 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 17 03:22:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	126689	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	213949	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	189559	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	79117	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	96843	45.731	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.460%
35) Dibromofluoromethane	5.373	113	73911	46.478	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.960%
50) Toluene-d8	8.635	98	260868	45.949	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.900%#
62) 4-Bromofluorobenzene	11.061	95	95077	46.645	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.280%
Target Compounds						
17) Carbon Disulfide	2.520	76	1589	0.398	ug/l	# 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121625\
Data File : VX048903.D
Acq On : 16 Dec 2025 21:16
Operator : JC/MD
Sample : Q3850-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Dec 17 03:30:25 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121625W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 17 03:22:16 2025
Response via : Initial Calibration

