

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121625\
 Data File : VX048908.D
 Acq On : 16 Dec 2025 23:00
 Operator : JC/MD
 Sample : Q3849-08DL 400X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-06-7-121125DL

Quant Time: Dec 17 03:31:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 17 03:22:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.537	168	109584	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	192084	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	171711	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	73115	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	86546	47.248	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.500%
35) Dibromofluoromethane	5.367	113	63899	44.756	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.520%
50) Toluene-d8	8.634	98	239492	46.985	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.980%
62) 4-Bromofluorobenzene	11.061	95	85082	46.493	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.980%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.337	101	8513	7.310	ug/l	98
27) cis-1,2-Dichloroethene	4.489	96	1006	0.611	ug/l	78
44) Trichloroethene	7.110	130	56665	38.257	ug/l	94
64) Tetrachloroethene	9.256	164	312	0.283	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121625\
Data File : VX048908.D
Acq On : 16 Dec 2025 23:00
Operator : JC/MD
Sample : Q3849-08DL 400X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-06-7-121125DL

Quant Time: Dec 17 03:31:56 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121625W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 17 03:22:16 2025
Response via : Initial Calibration

