



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121625\
Data File : VX048885.D
Acq On : 16 Dec 2025 13:01
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121625W.M
Title : SW846 8260
Last Update : Wed Dec 17 03:22:16 2025



AutoFind: Scans 1635, 1636, 1637; Background Corrected with Scan 1629

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.1	4679	PASS
75	95	30	60	52.1	11050	PASS
95	95	100	100	100.0	21216	PASS
96	95	5	9	6.0	1275	PASS
173	174	0.00	2	1.8	297	PASS
174	95	50	100	77.1	16356	PASS
175	174	5	9	7.2	1174	PASS
176	174	95	101	96.9	15845	PASS
177	176	5	9	6.4	1018	PASS