

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121718\
 Data File : VX006566.D
 Acq On : 17 Dec 2018 12:19
 Operator : JC/MD
 Sample : J6405-14
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB02-20181211

Quant Time: Dec 18 00:59:30 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 30 03:55:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	686787	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1068584	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	969517	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	457063	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	369992	52.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.54%	
35) Dibromofluoromethane	5.51	113	329792	47.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.34%	
50) Toluene-d8	8.71	98	1281391	50.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.84%	
62) 4-Bromofluorobenzene	11.14	95	465392	49.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.14%	
Target Compounds						
16) Acetone	2.45	43	13835	4.205	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121718\
Data File : VX006566.D
Acq On : 17 Dec 2018 12:19
Operator : JC/MD
Sample : J6405-14
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB02-20181211

Quant Time: Dec 18 00:59:30 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 30 03:55:33 2018
Response via : Initial Calibration

