

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121718\
 Data File : VX006572.D
 Acq On : 17 Dec 2018 14:59
 Operator : JC/MD
 Sample : J6405-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE108D1-20181210

Quant Time: Dec 18 01:17:40 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 30 03:55:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	683457	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1081583	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	982832	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	469861	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	376458	53.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.92%	
35) Dibromofluoromethane	5.50	113	329270	47.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.04%	
50) Toluene-d8	8.71	98	1292167	50.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.46%	
62) 4-Bromofluorobenzene	11.13	95	471092	49.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.16%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.60	96	2526	0.344	ug/l	73
44) Trichloroethene	7.21	130	315260	38.148	ug/l	97
49) 1,4-Dioxane	7.76	88	2818	13.820	ug/l #	70
64) Tetrachloroethene	9.34	164	10898	1.472	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121718\
Data File : VX006572.D
Acq On : 17 Dec 2018 14:59
Operator : JC/MD
Sample : J6405-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE108D1-20181210

Quant Time: Dec 18 01:17:40 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 30 03:55:33 2018
Response via : Initial Calibration

