

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121718\
 Data File : VX006578.D
 Acq On : 17 Dec 2018 17:39
 Operator : JC/MD
 Sample : J6408-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SA.PZ139I-20181208

Quant Time: Dec 18 01:28:30 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 30 03:55:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	666640	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1048825	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	968019	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	456233	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	367290	53.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.94%	
35) Dibromofluoromethane	5.50	113	320831	47.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.50%	
50) Toluene-d8	8.71	98	1260264	51.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.04%	
62) 4-Bromofluorobenzene	11.13	95	464860	49.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.88%	
Target Compounds						
44) Trichloroethene	7.21	130	16862	2.104	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121718\
Data File : VX006578.D
Acq On : 17 Dec 2018 17:39
Operator : JC/MD
Sample : J6408-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SA.PZ139I-20181208

Quant Time: Dec 18 01:28:30 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 30 03:55:33 2018
Response via : Initial Calibration

