

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121718\
 Data File : VX006579.D
 Acq On : 17 Dec 2018 18:06
 Operator : JC/MD
 Sample : J6408-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SA.MW130I-20181208

Quant Time: Dec 18 01:30:26 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 30 03:55:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	651097	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1050539	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	961288	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	460590	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	362995	54.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.22%	
35) Dibromofluoromethane	5.51	113	316435	46.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.04%	
50) Toluene-d8	8.71	98	1246918	50.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.80%	
62) 4-Bromofluorobenzene	11.13	95	459986	49.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.68%	
Target Compounds						
24) 1,1-Dichloroethane	3.70	63	9854	0.948	ug/l #	96
44) Trichloroethene	7.22	130	10945	1.364	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121718\
Data File : VX006579.D
Acq On : 17 Dec 2018 18:06
Operator : JC/MD
Sample : J6408-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SA.MW130I-20181208

Quant Time: Dec 18 01:30:26 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 30 03:55:33 2018
Response via : Initial Calibration

