

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121719\
 Data File : VX014052.D
 Acq On : 17 Dec 2019 15:17
 Operator : JC/SP
 Sample : VX1217WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1217WBL01

Quant Time: Dec 18 07:33:01 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 17 03:01:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	593364	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	912426	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	813423	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	362810	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.05	65	329941	49.06	ug/l	0.00
Spiked Amount			Recovery	=	98.12%	
35) Dibromofluoromethane	5.48	113	271670	48.98	ug/l	0.00
Spiked Amount			Recovery	=	97.96%	
50) Toluene-d8	8.71	98	1083687	50.18	ug/l	0.00
Spiked Amount			Recovery	=	100.36%	
62) 4-Bromofluorobenzene	11.13	95	363447	46.04	ug/l	0.00
Spiked Amount			Recovery	=	92.08%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121719\
Data File : VX014052.D
Acq On : 17 Dec 2019 15:17
Operator : JC/SP
Sample : VX1217WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1217WBL01

Quant Time: Dec 18 07:33:01 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 17 03:01:07 2019
Response via : Initial Calibration

