

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121719\
 Data File : VX014056.D
 Acq On : 17 Dec 2019 16:59
 Operator : JC/SP
 Sample : K6251-11
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW201D-20191207

Quant Time: Dec 18 07:42:55 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 17 03:01:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	563699	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	865170	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	772438	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	324040	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.05	65	319054	49.94	ug/l	0.00
Spiked Amount			Recovery	=	99.88%	
35) Dibromofluoromethane	5.49	113	260071	49.45	ug/l	0.00
Spiked Amount			Recovery	=	98.90%	
50) Toluene-d8	8.71	98	1023469	49.98	ug/l	0.00
Spiked Amount			Recovery	=	99.96%	
62) 4-Bromofluorobenzene	11.14	95	338232	45.19	ug/l	0.00
Spiked Amount			Recovery	=	90.38%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Dichlorodifluoromethane	1.19	85	31725	6.291	ug/l	98
7) Trichlorofluoromethane	1.92	101	870478	98.468	ug/l	99
9) 1,1,2-Trichlorotrifluoroet	2.37	101	11809	2.216	ug/l	95
12) 1,1-Dichloroethene	2.37	96	165218	30.469	ug/l	99
24) 1,1-Dichloroethane	3.69	63	33207	3.094	ug/l	99
27) cis-1,2-Dichloroethene	4.59	96	15427	2.283	ug/l	87
30) Chloroform	5.20	83	36746	3.611	ug/l	97
32) 1,1,1-Trichloroethane	5.48	97	42968	4.955	ug/l	98
44) Trichloroethene	7.21	130	413051	61.224	ug/l	98
55) 1,1,2-Trichloroethane	9.21	97	2503	0.405	ug/l #	65
64) Tetrachloroethene	9.33	164	386301	56.517	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121719\
Data File : VX014056.D
Acq On : 17 Dec 2019 16:59
Operator : JC/SP
Sample : K6251-11
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW201D-20191207

Quant Time: Dec 18 07:42:55 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 17 03:01:07 2019
Response via : Initial Calibration

