

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121719\
 Data File : VX014063.D
 Acq On : 17 Dec 2019 19:42
 Operator : JC/SP
 Sample : K6251-21
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE132D1-20191209

Quant Time: Dec 18 07:56:58 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 17 03:01:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	553916	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	863581	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	769575	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	318536	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.05	65	313545	49.94	ug/l	0.00
Spiked Amount			Recovery	=	99.88%	
35) Dibromofluoromethane	5.49	113	257355	49.02	ug/l	0.00
Spiked Amount			Recovery	=	98.04%	
50) Toluene-d8	8.71	98	1010705	49.45	ug/l	0.00
Spiked Amount			Recovery	=	98.90%	
62) 4-Bromofluorobenzene	11.13	95	331162	44.33	ug/l	0.00
Spiked Amount			Recovery	=	88.66%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
9) 1,1,2-Trichlorotrifluoroet	2.37	101	30454	5.816	ug/l	96
12) 1,1-Dichloroethene	2.36	96	7024	1.318	ug/l	83
19) Methyl tert-butyl Ether	3.19	73	22274	1.272	ug/l	91
22) Diisopropyl ether	3.84	45	13012	0.686	ug/l #	73
27) cis-1,2-Dichloroethene	4.59	96	13998	2.108	ug/l	84
30) Chloroform	5.20	83	20097	2.010	ug/l	97
44) Trichloroethene	7.20	130	739683	109.840	ug/l	96
64) Tetrachloroethene	9.33	164	19905	2.923	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121719\
Data File : VX014063.D
Acq On : 17 Dec 2019 19:42
Operator : JC/SP
Sample : K6251-21
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE132D1-20191209

Quant Time: Dec 18 07:56:58 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 17 03:01:07 2019
Response via : Initial Calibration

