

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121720\
 Data File : VX020079.D
 Acq On : 16 Dec 2020 20:36
 Operator : JC/MD
 Sample : L5097-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ID039-39MWE10-120720

Quant Time: Dec 17 00:31:09 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	284715	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.830	114	478950	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.098	117	447744	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	190393	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.031	65	187528	50.11	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	100.22%
35) Dibromofluoromethane	5.464	113	151245	48.86	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.72%
50) Toluene-d8	8.695	98	583621	49.39	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.78%
62) 4-Bromofluorobenzene	11.122	95	206417	45.89	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	91.78%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121720\
Data File : VX020079.D
Acq On : 16 Dec 2020 20:36
Operator : JC/MD
Sample : L5097-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ID039-39MWE10-120720

Quant Time: Dec 17 00:31:09 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration

