

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121720\
 Data File : VX020105.D
 Acq On : 17 Dec 2020 07:27
 Operator : JC/MD
 Sample : L5116-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-2

Quant Time: Dec 17 07:54:26 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	291985	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.830	114	494331	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.092	117	452496	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	192593	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.031	65	191646	49.94	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	99.88%		
35) Dibromofluoromethane	5.464	113	153308	47.99	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	95.98%		
50) Toluene-d8	8.696	98	596351	48.89	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	97.78%		
62) 4-Bromofluorobenzene	11.122	95	207811	44.76	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	89.52%		
Target Compounds						
16) Acetone	2.416	43	2036	1.41	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121720\
Data File : VX020105.D
Acq On : 17 Dec 2020 07:27
Operator : JC/MD
Sample : L5116-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-2

Quant Time: Dec 17 07:54:26 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration

