

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121720\
 Data File : VX020065.D
 Acq On : 16 Dec 2020 15:06
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 17 00:28:12 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : Sw846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	78	0.00
2 T	Dichlorodifluoromethane	50.000	45.052	9.9	66	0.00
3 P	Chloromethane	50.000	40.703	18.6	65	0.00
4 C	Vinyl Chloride	50.000	47.452	5.1#	72	0.00
5 T	Bromomethane	50.000	40.041	19.9	61	0.00
6 T	Chloroethane	50.000	52.914	-5.8	79	0.00
7 T	Trichlorofluoromethane	50.000	49.525	1.0	76	0.00
8 T	Diethyl Ether	50.000	50.709	-1.4	80	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	51.687	-3.4	79	0.00
10 T	Methyl Iodide	50.000	37.766	24.5	58	0.00
11 T	Tert butyl alcohol	250.000	232.980	6.8	74	0.00
12 CM	1,1-Dichloroethene	50.000	48.370	3.3#	74	0.00
13 T	Acrolein	250.000	184.653	26.1#	60	0.00
14 T	Allyl chloride	50.000	51.390	-2.8	80	0.00
15 T	Acrylonitrile	250.000	264.681	-5.9	83	0.00
16 T	Acetone	250.000	275.060	-10.0	85	0.00
17 T	Carbon Disulfide	50.000	40.927	18.1	64	0.00
18 T	Methyl Acetate	50.000	57.298	-14.6	88	0.00
19 T	Methyl tert-butyl Ether	50.000	52.940	-5.9	81	0.00
20 T	Methylene Chloride	50.000	51.884	-3.8	78	0.00
21 T	trans-1,2-Dichloroethene	50.000	49.225	1.5	75	0.00
22 T	Diisopropyl ether	50.000	55.523	-11.0	85	0.00
23 T	Vinyl Acetate	250.000	270.928	-8.4	82	0.00
24 P	1,1-Dichloroethane	50.000	53.166	-6.3	82	0.00
25 T	2-Butanone	250.000	272.212	-8.9	85	0.00
26 T	2,2-Dichloropropane	50.000	52.303	-4.6	80	0.00
27 T	cis-1,2-Dichloroethene	50.000	50.957	-1.9	80	0.00
28 T	Bromochloromethane	50.000	50.939	-1.9	79	0.00
29 T	Tetrahydrofuran	250.000	260.048	-4.0	81	0.00
30 C	Chloroform	50.000	53.636	-7.3#	83	0.00
31 T	Cyclohexane	50.000	49.889	0.2	75	0.00
32 T	1,1,1-Trichloroethane	50.000	51.867	-3.7	79	0.00
33 S	1,2-Dichloroethane-d4	50.000	47.625	4.8	78	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	76	0.00
35 S	Dibromofluoromethane	50.000	50.058	-0.1	79	0.00
36 T	1,1-Dichloropropene	50.000	50.405	-0.8	77	0.00
37 T	Ethyl Acetate	50.000	52.202	-4.4	82	0.00
38 T	Carbon Tetrachloride	50.000	51.776	-3.6	77	0.00
39 T	Methylcyclohexane	50.000	50.854	-1.7	75	0.00
40 TM	Benzene	50.000	52.284	-4.6	80	0.00
41 T	Methacrylonitrile	50.000	54.284	-8.6	83	0.00
42 TM	1,2-Dichloroethane	50.000	53.044	-6.1	80	0.00
43 T	Isopropyl Acetate	50.000	52.393	-4.8	81	0.00
44 TM	Trichloroethene	50.000	50.663	-1.3	77	0.00
45 C	1,2-Dichloropropane	50.000	54.839	-9.7#	83	0.00
46 T	Dibromomethane	50.000	53.059	-6.1	80	0.00
47 T	Bromodichloromethane	50.000	54.468	-8.9	81	0.00
48 T	Methyl methacrylate	50.000	53.319	-6.6	81	0.00
49 T	1,4-Dioxane	1000.000	995.743	0.4	76	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121720\
 Data File : VX020065.D
 Acq On : 16 Dec 2020 15:06
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 17 00:28:12 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : Sw846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)

50 S	Toluene-d8	50.000	49.426	1.1	76	0.00
51 T	4-Methyl-2-Pentanone	250.000	274.350	-9.7	83	0.00
52 CM	Toluene	50.000	52.264	-4.5#	79	0.00
53 T	t-1,3-Dichloropropene	50.000	53.528	-7.1	78	0.00
54 T	cis-1,3-Dichloropropene	50.000	54.250	-8.5	79	0.00
55 T	1,1,2-Trichloroethane	50.000	55.401	-10.8	84	0.00
56 T	Ethyl methacrylate	50.000	54.344	-8.7	80	0.00
57 T	1,3-Dichloropropane	50.000	53.682	-7.4	82	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	276.759	-10.7	84	0.00
59 T	2-Hexanone	250.000	273.065	-9.2	82	0.00
60 T	Dibromochloromethane	50.000	53.538	-7.1	78	0.00
61 T	1,2-Dibromoethane	50.000	52.241	-4.5	80	0.00
62 S	4-Bromofluorobenzene	50.000	50.112	-0.2	77	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	78	0.00
64 T	Tetrachloroethene	50.000	53.437	-6.9	83	0.00
65 PM	Chlorobenzene	50.000	50.930	-1.9	79	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	52.811	-5.6	80	0.00
67 C	Ethyl Benzene	50.000	51.811	-3.6#	79	0.00
68 T	m/p-Xylenes	100.000	102.739	-2.7	78	0.00
69 T	o-Xylene	50.000	51.237	-2.5	78	0.00
70 T	Styrene	50.000	53.019	-6.0	79	0.00
71 P	Bromoform	50.000	51.330	-2.7	75	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	76	0.00
73 T	Isopropylbenzene	50.000	52.669	-5.3	80	0.00
74 T	N-aryl acetate	50.000	51.764	-3.5	77	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	51.877	-3.8	79	0.00
76 T	1,2,3-Trichloropropane	50.000	52.285	-4.6	80	0.00
77 T	Bromobenzene	50.000	50.901	-1.8	79	0.00
78 T	n-propylbenzene	50.000	53.706	-7.4	80	0.00
79 T	2-Chlorotoluene	50.000	52.213	-4.4	80	0.00
80 T	1,3,5-Trimethylbenzene	50.000	53.076	-6.2	78	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	51.115	-2.2	75	0.00
82 T	4-Chlorotoluene	50.000	52.490	-5.0	80	0.00
83 T	tert-Butylbenzene	50.000	52.468	-4.9	78	0.00
84 T	1,2,4-Trimethylbenzene	50.000	53.347	-6.7	78	0.00
85 T	sec-Butylbenzene	50.000	54.351	-8.7	80	0.00
86 T	p-Isopropyltoluene	50.000	55.282	-10.6	81	0.00
87 T	1,3-Dichlorobenzene	50.000	51.192	-2.4	78	0.00
88 T	1,4-Dichlorobenzene	50.000	50.382	-0.8	77	0.00
89 T	n-Butylbenzene	50.000	56.054	-12.1	82	0.00
90 T	Hexachloroethane	50.000	52.511	-5.0	79	0.00
91 T	1,2-Dichlorobenzene	50.000	51.859	-3.7	78	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	51.175	-2.3	79	0.00
93 T	1,2,4-Trichlorobenzene	50.000	54.233	-8.5	80	0.00
94 T	Hexachlorobutadiene	50.000	56.865	-13.7	89	0.00
95 T	Naphthalene	50.000	54.913	-9.8	81	0.00
96 T	1,2,3-Trichlorobenzene	50.000	55.294	-10.6	82	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121720\
Data File : VX020065.D
Acq On : 16 Dec 2020 15:06
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Dec 17 00:28:12 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420w.M
Quant Title : SW846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 6