

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121721\
 Data File : VX025849.D
 Acq On : 16 Dec 2021 12:21
 Operator : JC/MD
 Sample : VX1217MBL01
 Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1217MBL01

Quant Time: Dec 17 02:01:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 14 07:02:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	143876	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	246389	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	219802	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	92294	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	95938	46.344	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.680%
35) Dibromofluoromethane	5.379	113	79318	46.160	ug/l	-0.01
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.320%
50) Toluene-d8	8.647	98	291670	46.587	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.180%
62) 4-Bromofluorobenzene	11.079	95	97916	41.171	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	82.340%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121721\
Data File : VX025849.D
Acq On : 16 Dec 2021 12:21
Operator : JC/MD
Sample : VX1217MBL01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1217MBL01

Quant Time: Dec 17 02:01:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 14 07:02:11 2021
Response via : Initial Calibration

