

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121721\
 Data File : VX025856.D
 Acq On : 16 Dec 2021 15:31
 Operator : JC/MD
 Sample : M5084-05RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 C6RE

Quant Time: Dec 17 02:31:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 14 07:02:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	150393	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	255219	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	230395	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	102113	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	97562	45.086	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.180%
35) Dibromofluoromethane	5.391	113	65946	37.050	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	74.100%#
50) Toluene-d8	8.653	98	298222	45.985	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.980%#
62) 4-Bromofluorobenzene	11.079	95	105338	42.760	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	85.520%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	53570	39.571	ug/l	97
25) 2-Butanone	4.562	43	16787	11.534	ug/l	95
30) Chloroform	5.099	83	9341	2.646	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121721\
Data File : VX025856.D
Acq On : 16 Dec 2021 15:31
Operator : JC/MD
Sample : M5084-05RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
C6RE

Quant Time: Dec 17 02:31:19 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 14 07:02:11 2021
Response via : Initial Calibration

