

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121721\
 Data File : VX025857.D
 Acq On : 16 Dec 2021 15:54
 Operator : JC/MD
 Sample : M5084-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 C7

Quant Time: Dec 17 02:31:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 14 07:02:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	138457	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	239784	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	214819	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	95837	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	94297	47.334	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.660%
35) Dibromofluoromethane	5.391	113	75929	45.404	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.800%
50) Toluene-d8	8.653	98	279864	45.932	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.860%#
62) 4-Bromofluorobenzene	11.085	95	99970	43.193	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.380%
Target Compounds						
					Qvalue	
73) Isopropylbenzene	10.963	105	9108	1.309	ug/l	97
83) tert-Butylbenzene	11.719	119	3901	0.690	ug/l	84
85) sec-Butylbenzene	11.890	105	9731	1.432	ug/l #	70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121721\
Data File : VX025857.D
Acq On : 16 Dec 2021 15:54
Operator : JC/MD
Sample : M5084-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
C7

Quant Time: Dec 17 02:31:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 14 07:02:11 2021
Response via : Initial Calibration

