

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121721\
 Data File : VX025859.D
 Acq On : 16 Dec 2021 16:42
 Operator : JC/MD
 Sample : PB141455TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB141455TB

Quant Time: Dec 17 02:32:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 14 07:02:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	147431	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	257576	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	242612	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	100549	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	98102	46.247	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.500%
35) Dibromofluoromethane	5.391	113	80568	44.851	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.700%
50) Toluene-d8	8.653	98	308891	47.194	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.380%
62) 4-Bromofluorobenzene	11.085	95	108120	43.487	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.980%
Target Compounds						
25) 2-Butanone	4.580	43	5116	3.586	ug/l	96
43) Isopropyl Acetate	6.348	43	13173	3.021	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121721\
Data File : VX025859.D
Acq On : 16 Dec 2021 16:42
Operator : JC/MD
Sample : PB141455TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB141455TB

Quant Time: Dec 17 02:32:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 14 07:02:11 2021
Response via : Initial Calibration

