

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121721\
 Data File : VX025860.D
 Acq On : 16 Dec 2021 17:05
 Operator : JC/MD
 Sample : PB141455ZHE#01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB141455ZHE#01

Quant Time: Dec 17 02:32:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 14 07:02:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	140299	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	245805	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	229014	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	93167	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	94456	46.792	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	93.580%		
35) Dibromofluoromethane	5.391	113	75838	44.239	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	88.480%		
50) Toluene-d8	8.652	98	294298	47.118	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	94.240%		
62) 4-Bromofluorobenzene	11.085	95	101676	42.854	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	85.700%		
Target Compounds						
25) 2-Butanone	4.574	43	4717	3.474	ug/l	92
43) Isopropyl Acetate	6.354	43	13368	3.213	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121721\
Data File : VX025860.D
Acq On : 16 Dec 2021 17:05
Operator : JC/MD
Sample : PB141455ZHE#01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB141455ZHE#01

Quant Time: Dec 17 02:32:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 14 07:02:11 2021
Response via : Initial Calibration

