

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121721\
 Data File : VX025863.D
 Acq On : 16 Dec 2021 18:16
 Operator : JC/MD
 Sample : PB141455ZHE#04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB141455ZHE#04

Quant Time: Dec 17 02:34:36 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 14 07:02:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	142221	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	243709	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	230883	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	91357	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	96399	47.109	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.220%
35) Dibromofluoromethane	5.391	113	77613	45.664	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.320%
50) Toluene-d8	8.653	98	298481	48.199	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.400%
62) 4-Bromofluorobenzene	11.085	95	101913	43.323	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.640%
Target Compounds						
					Qvalue	
18) Methyl Acetate	2.715	43	1034	0.332	ug/l #	81
20) Methylene Chloride	2.788	84	6084	1.111	ug/l	89
25) 2-Butanone	4.574	43	2225	1.617	ug/l	99
43) Isopropyl Acetate	6.348	43	13175	3.193	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121721\
Data File : VX025863.D
Acq On : 16 Dec 2021 18:16
Operator : JC/MD
Sample : PB141455ZHE#04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB141455ZHE#04

Quant Time: Dec 17 02:34:36 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 14 07:02:11 2021
Response via : Initial Calibration

