

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121721\
 Data File : VX025864.D
 Acq On : 16 Dec 2021 18:40
 Operator : JC/MD
 Sample : PB141455ZHE#05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB141455ZHE#05

Quant Time: Dec 17 02:35:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 14 07:02:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	139046	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	244998	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	233962	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	98893	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	95651	47.811	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.620%
35) Dibromofluoromethane	5.391	113	77996	45.648	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.300%
50) Toluene-d8	8.653	98	295989	47.545	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.080%
62) 4-Bromofluorobenzene	11.085	95	103785	43.887	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.780%
Target Compounds						
					Qvalue	
20) Methylene Chloride	2.794	84	6313	1.335	ug/l	97
43) Isopropyl Acetate	6.354	43	13816	3.331	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121721\
Data File : VX025864.D
Acq On : 16 Dec 2021 18:40
Operator : JC/MD
Sample : PB141455ZHE#05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB141455ZHE#05

Quant Time: Dec 17 02:35:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 14 07:02:11 2021
Response via : Initial Calibration

