

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121721\
 Data File : VX025867.D
 Acq On : 16 Dec 2021 19:51
 Operator : JC/MD
 Sample : M5051-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 P-3

Quant Time: Dec 17 02:36:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 14 07:02:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	131535	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	229042	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	221349	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	93388	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	91981	48.602	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.200%
35) Dibromofluoromethane	5.391	113	73843	46.228	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.460%
50) Toluene-d8	8.653	98	280234	48.150	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.300%
62) 4-Bromofluorobenzene	11.085	95	99269	44.902	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.800%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	18128	1.282	ug/l	97
20) Methylene Chloride	2.788	84	10049	3.981	ug/l	96
25) 2-Butanone	4.574	43	2487	1.954	ug/l	98
43) Isopropyl Acetate	6.349	43	11988	3.092	ug/l	97
69) o-Xylene	10.647	106	3252	1.089	ug/l	98
80) 1,3,5-Trimethylbenzene	11.451	105	9727	1.730	ug/l	95
84) 1,2,4-Trimethylbenzene	11.756	105	7035	1.253	ug/l	97
95) Naphthalene	13.780	128	4714	3.115	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121721\
Data File : VX025867.D
Acq On : 16 Dec 2021 19:51
Operator : JC/MD
Sample : M5051-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
P-3

Quant Time: Dec 17 02:36:46 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 14 07:02:11 2021
Response via : Initial Calibration

