

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121721\
 Data File : VX025868.D
 Acq On : 16 Dec 2021 20:14
 Operator : JC/MD
 Sample : M5051-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 P-2

Quant Time: Dec 17 02:37:20 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 14 07:02:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.555	168	131383	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	229720	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	222230	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	93100	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	91704	48.511	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.020%
35) Dibromofluoromethane	5.391	113	74803	46.691	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.380%
50) Toluene-d8	8.652	98	284164	48.681	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.360%
62) 4-Bromofluorobenzene	11.085	95	98461	44.405	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.800%
Target Compounds						
					Qvalue	
20) Methylene Chloride	2.788	84	9342	3.529	ug/l	93
43) Isopropyl Acetate	6.354	43	11625	2.989	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121721\
Data File : VX025868.D
Acq On : 16 Dec 2021 20:14
Operator : JC/MD
Sample : M5051-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
P-2

Quant Time: Dec 17 02:37:20 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 14 07:02:11 2021
Response via : Initial Calibration

