

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121724\
 Data File : VX044360.D
 Acq On : 17 Dec 2024 18:37
 Operator : JC/MD
 Sample : P5263-01 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DCZF0

Quant Time: Dec 18 04:42:54 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML120524WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Dec 18 04:40:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.757	114	236173	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	202347	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	78934	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	70740	32.731	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	65.460%
7) Chloroethane-d5	1.648	69	65941	39.386	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	78.780%
11) 1,1-Dichloroethene-d2	2.294	65	31887	34.498	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	69.000%
21) 2-Butanone-d5	4.458	46	119287	90.148	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	90.150%
24) Chloroform-d	5.056	84	164380	40.058	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	80.120%
26) 1,2-Dichloroethane-d4	5.952	65	113708	39.183	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	78.360%
32) Benzene-d6	5.970	84	300342	42.957	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.920%
36) 1,2-Dichloropropane-d6	7.305	67	100044	46.082	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	92.160%
41) Toluene-d8	8.647	98	255661	39.547	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	79.100%#
43) trans-1,3-Dichloroprop...	8.952	79	37509	35.953	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.900%
47) 2-Hexanone-d5	9.384	63	81009	95.700	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	95.700%
56) 1,1,2,2-Tetrachloroeth...	11.189	84	131718	45.248	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	90.500%
66) 1,2-Dichlorobenzene-d4	12.317	152	84254	45.812	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.620%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121724\
Data File : VX044360.D
Acq On : 17 Dec 2024 18:37
Operator : JC/MD
Sample : P5263-01 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DCZF0

Quant Time: Dec 18 04:42:54 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM120524WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Dec 18 04:40:20 2024
Response via : Initial Calibration

