

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121724\
 Data File : VX044350.D
 Acq On : 17 Dec 2024 14:42
 Operator : JC/MD
 Sample : P5235-11 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 C0AW1

Quant Time: Dec 18 03:08:33 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML120524WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Dec 18 03:03:46 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.757	114	252649	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	218757	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	87370	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	76814	33.223	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	66.440%
7) Chloroethane-d5	1.648	69	74514	41.605	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	83.200%
11) 1,1-Dichloroethene-d2	2.294	65	33823	34.207	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	68.420%
21) 2-Butanone-d5	4.465	46	124005	87.602	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	87.600%
24) Chloroform-d	5.056	84	167585	38.176	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	76.360%
26) 1,2-Dichloroethane-d4	5.952	65	116406	37.497	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	75.000%
32) Benzene-d6	5.971	84	304616	40.300	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	80.600%
36) 1,2-Dichloropropane-d6	7.306	67	98807	42.098	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	84.200%
41) Toluene-d8	8.647	98	265044	37.923	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	75.840%#
43) trans-1,3-Dichloroprop...	8.952	79	39600	35.110	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	70.220%
47) 2-Hexanone-d5	9.385	63	81446	88.999	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	89.000%
56) 1,1,2,2-Tetrachloroeth...	11.189	84	131624	41.824	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	83.640%
66) 1,2-Dichlorobenzene-d4	12.317	152	88180	43.317	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	86.640%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121724\
Data File : VX044350.D
Acq On : 17 Dec 2024 14:42
Operator : JC/MD
Sample : P5235-11 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
C0AW1

Quant Time: Dec 18 03:08:33 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM120524WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Dec 18 03:03:46 2024
Response via : Initial Calibration

