

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121818\
 Data File : VX006607.D
 Acq On : 18 Dec 2018 19:25
 Operator : JC/MD
 Sample : J6405-10DL 50X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE108D2-20181210DL

Quant Time: Dec 19 07:06:10 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 18 13:23:17 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	675169	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1054189	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	965657	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	467964	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	361147	49.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.26%	
35) Dibromofluoromethane	5.50	113	313928	47.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.18%	
50) Toluene-d8	8.71	98	1243233	49.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.94%	
62) 4-Bromofluorobenzene	11.13	95	459225	49.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.78%	
Target Compounds						
44) Trichloroethene	7.21	130	424786	52.882	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121818\
Data File : VX006607.D
Acq On : 18 Dec 2018 19:25
Operator : JC/MD
Sample : J6405-10DL 50X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE108D2-20181210DL

Quant Time: Dec 19 07:06:10 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 18 13:23:17 2018
Response via : Initial Calibration

