

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121818\
 Data File : VX006619.D
 Acq On : 19 Dec 2018 01:38
 Operator : JC/MD
 Sample : J6408-12
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SA.1571-20181208

Quant Time: Dec 19 07:25:33 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 18 13:23:17 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	657820	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1058380	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	965899	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	460223	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	363487	51.27	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.54%	
35) Dibromofluoromethane	5.50	113	315538	47.14	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.28%	
50) Toluene-d8	8.71	98	1261600	50.00	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.00%	
62) 4-Bromofluorobenzene	11.13	95	454071	49.13	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.26%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121818\
Data File : VX006619.D
Acq On : 19 Dec 2018 01:38
Operator : JC/MD
Sample : J6408-12
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SA.157I-20181208

Quant Time: Dec 19 07:25:33 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 18 13:23:17 2018
Response via : Initial Calibration

