

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121818\
 Data File : VX006620.D
 Acq On : 19 Dec 2018 02:05
 Operator : JC/MD
 Sample : J6408-13
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SA.PZ182I-20181208

Quant Time: Dec 19 07:27:13 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 18 13:23:17 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	622022	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	987220	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	922383	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	441232	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	349347	52.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.22%	
35) Dibromofluoromethane	5.51	113	296679	47.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.04%	
50) Toluene-d8	8.71	98	1185605	50.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.74%	
62) 4-Bromofluorobenzene	11.13	95	426633	49.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.98%	
Target Compounds						
24) 1,1-Dichloroethane	3.69	63	12576	1.231	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121818\
Data File : VX006620.D
Acq On : 19 Dec 2018 02:05
Operator : JC/MD
Sample : J6408-13
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SA.PZ182I-20181208

Quant Time: Dec 19 07:27:13 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 18 13:23:17 2018
Response via : Initial Calibration

