

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121818\
 Data File : VX006622.D
 Acq On : 19 Dec 2018 02:58
 Operator : JC/MD
 Sample : J6408-16
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SA.PZ134I-20181208

Quant Time: Dec 19 07:30:05 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 18 13:23:17 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	636757	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1013519	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	930780	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	450512	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	354511	51.66	ug/l	0.00
Spiked Amount			Recovery	=	103.32%	
35) Dibromofluoromethane	5.51	113	300100	46.82	ug/l	0.00
Spiked Amount			Recovery	=	93.64%	
50) Toluene-d8	8.71	98	1213564	50.22	ug/l	0.00
Spiked Amount			Recovery	=	100.44%	
62) 4-Bromofluorobenzene	11.14	95	447567	50.57	ug/l	0.00
Spiked Amount			Recovery	=	101.14%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121818\
Data File : VX006622.D
Acq On : 19 Dec 2018 02:58
Operator : JC/MD
Sample : J6408-16
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SA.PZ134I-20181208

Quant Time: Dec 19 07:30:05 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 18 13:23:17 2018
Response via : Initial Calibration

