

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121818\
 Data File : VX006629.D
 Acq On : 19 Dec 2018 06:05
 Operator : JC/MD
 Sample : J6449-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-17A_R1

Quant Time: Dec 19 07:43:43 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 18 13:23:17 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	623359	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	998192	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	922223	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	443642	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	353374	52.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.20%	
35) Dibromofluoromethane	5.50	113	296723	47.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.02%	
50) Toluene-d8	8.71	98	1194375	50.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.38%	
62) 4-Bromofluorobenzene	11.13	95	439877	50.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.94%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.60	96	54633	8.024	ug/l	89
44) Trichloroethene	7.22	130	8952	1.177	ug/l	86
64) Tetrachloroethene	9.34	164	9289	1.272	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121818\
Data File : VX006629.D
Acq On : 19 Dec 2018 06:05
Operator : JC/MD
Sample : J6449-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
MW-17A_R1

Quant Time: Dec 19 07:43:43 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 18 13:23:17 2018
Response via : Initial Calibration

