

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\WX121818\
 Data File : VX006633.D
 Acq On : 19 Dec 2018 07:51
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 19 10:39:48 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 19 00:58:54 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	94	0.00
2 T	Dichlorodifluoromethane	0.457	0.436	4.6	93	0.00
3 P	Chloromethane	0.493	0.482	2.2	96	0.00
4 C	Vinyl Chloride	0.515	0.514	0.2#	98	0.00
5 T	Bromomethane	0.560	0.515	8.0	99	0.00
6 T	Chloroethane	0.356	0.331	7.0	97	0.00
7 T	Trichlorofluoromethane	0.802	0.796	0.7	95	0.00
8 T	Diethyl Ether	0.319	0.306	4.1	96	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.471	0.446	5.3	92	0.00
10 T	Methyl Iodide	0.674	0.732	-8.6	96	0.00
11 T	Tert butyl alcohol	0.120	0.123	-2.5	97	0.00
12 CM	1,1-Dichloroethene	0.447	0.438	2.0#	97	0.00
13 T	Acrolein	0.081	0.070	13.6	84	0.00
14 T	Allyl chloride	0.752	0.728	3.2	93	0.00
15 T	Acrylonitrile	0.267	0.274	-2.6	97	0.00
16 T	Acetone	0.248	0.227	8.5	90	0.00
17 T	Carbon Disulfide	1.182	1.110	6.1	93	0.00
18 T	Methyl Acetate	0.728	0.748	-2.7	99	0.00
19 T	Methyl tert-butyl Ether	1.483	1.520	-2.5	98	0.00
20 T	Methylene Chloride	0.512	0.495	3.3	98	0.00
21 T	trans-1,2-Dichloroethene	0.495	0.470	5.1	96	0.00
22 T	Diisopropyl ether	1.376	1.402	-1.9	97	0.00
23 T	Vinyl Acetate	1.125	1.126	-0.1	94	0.00
24 P	1,1-Dichloroethane	0.821	0.792	3.5	97	0.00
25 T	2-Butanone	0.348	0.343	1.4	94	0.00
26 T	2,2-Dichloropropane	0.634	0.418	34.1#	63	0.00
27 T	cis-1,2-Dichloroethene	0.546	0.545	0.2	95	0.00
28 T	Bromochloromethane	0.378	0.366	3.2	97	0.00
29 T	Tetrahydrofuran	0.229	0.233	-1.7	97	0.00
30 C	Chloroform	0.850	0.858	-0.9#	97	0.00
31 T	Cyclohexane	0.746	0.719	3.6	93	0.00
32 T	1,1,1-Trichloroethane	0.718	0.738	-2.8	95	0.00
33 S	1,2-Dichloroethane-d4	0.539	0.539	0.0	94	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	96	0.00
35 S	Dibromofluoromethane	0.316	0.310	1.9	94	0.00
36 T	1,1-Dichloropropene	0.414	0.400	3.4	95	0.00
37 T	Ethyl Acetate	0.431	0.420	2.6	98	0.00
38 T	Carbon Tetrachloride	0.402	0.397	1.2	95	0.00
39 T	Methylcyclohexane	0.545	0.499	8.4	92	0.00
40 TM	Benzene	1.259	1.226	2.6	96	0.00
41 T	Methacrylonitrile	0.229	0.233	-1.7	95	0.00
42 TM	1,2-Dichloroethane	0.429	0.407	5.1	97	0.00
43 T	Isopropyl Acetate	0.656	0.671	-2.3	97	0.00
44 TM	Trichloroethene	0.381	0.359	5.8	96	0.00
45 C	1,2-Dichloropropane	0.306	0.305	0.3#	101	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\WX121818\
 Data File : VX006633.D
 Acq On : 19 Dec 2018 07:51
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 19 10:39:48 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 19 00:58:54 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.225	0.218	3.1	95	0.00
47 T	Bromodichloromethane	0.374	0.374	0.0	95	0.00
48 T	Methyl methacrylate	0.334	0.341	-2.1	96	0.00
49 T	1,4-Dioxane	0.009	0.009	0.0	96	0.00
50 S	Toluene-d8	1.192	1.185	0.6	96	0.00
51 T	4-Methyl-2-Pentanone	0.417	0.434	-4.1	98	0.00
52 CM	Toluene	0.813	0.802	1.4#	96	0.00
53 T	t-1,3-Dichloropropene	0.401	0.391	2.5	88	0.00
54 T	cis-1,3-Dichloropropene	0.440	0.437	0.7	90	0.00
55 T	1,1,2-Trichloroethane	0.333	0.333	0.0	98	0.00
56 T	Ethyl methacrylate	0.470	0.494	-5.1	96	0.00
57 T	1,3-Dichloropropane	0.536	0.523	2.4	97	0.00
58 T	2-Chloroethyl Vinyl ether	0.250	0.246	1.6	95	0.00
59 T	2-Hexanone	0.319	0.328	-2.8	96	0.00
60 T	Dibromochloromethane	0.307	0.327	-6.5	96	0.00
61 T	1,2-Dibromoethane	0.354	0.360	-1.7	97	0.00
62 S	4-Bromofluorobenzene	0.437	0.441	-0.9	95	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	95	0.00
64 T	Tetrachloroethene	0.396	0.391	1.3	100	0.00
65 PM	Chlorobenzene	1.058	1.036	2.1	97	0.00
66 T	1,1,1,2-Tetrachloroethane	0.330	0.347	-5.2	96	0.00
67 C	Ethyl Benzene	1.729	1.704	1.4#	97	0.00
68 T	m/p-Xylenes	0.682	0.682	0.0	96	0.00
69 T	o-Xylene	0.674	0.675	-0.1	96	0.00
70 T	Styrene	1.066	1.092	-2.4	96	0.00
71 P	Bromoform	0.251	0.267	-6.4	95	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.00
73 T	Isopropylbenzene	3.421	3.373	1.4	95	0.00
74 T	N-amyl acetate	1.224	1.290	-5.4	98	0.00
75 P	1,1,2,2-Tetrachloroethane	1.017	1.033	-1.6	97	0.00
76 T	1,2,3-Trichloropropane	0.929	1.130	-21.6#	111	0.00
77 T	Bromobenzene	0.925	0.916	1.0	95	0.00
78 T	n-propylbenzene	3.767	3.750	0.5	95	0.00
79 T	2-Chlorotoluene	2.283	2.242	1.8	96	0.00
80 T	1,3,5-Trimethylbenzene	2.817	2.793	0.9	95	0.00
81 T	trans-1,4-Dichloro-2-butene	0.259	0.228	12.0	79	0.00
82 T	4-Chlorotoluene	2.618	2.557	2.3	94	0.00
83 T	tert-Butylbenzene	2.820	2.866	-1.6	95	0.00
84 T	1,2,4-Trimethylbenzene	2.963	2.906	1.9	96	0.00
85 T	sec-Butylbenzene	3.457	3.399	1.7	94	0.00
86 T	p-Isopropyltoluene	3.039	3.027	0.4	94	0.00
87 T	1,3-Dichlorobenzene	1.667	1.655	0.7	96	0.00
88 T	1,4-Dichlorobenzene	1.709	1.645	3.7	94	0.00
89 T	n-Butylbenzene	2.586	2.532	2.1	92	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121818\
Data File : VX006633.D
Acq On : 19 Dec 2018 07:51
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Dec 19 10:39:48 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 19 00:58:54 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.414	0.438	-5.8	95	0.00
91 T	1,2-Dichlorobenzene	1.667	1.620	2.8	95	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.210	0.211	-0.5	93	0.00
93 T	1,2,4-Trichlorobenzene	1.150	1.132	1.6	94	0.00
94 T	Hexachlorobutadiene	0.542	0.495	8.7	91	0.00
95 T	Naphthalene	3.671	3.630	1.1	95	0.00
96 T	1,2,3-Trichlorobenzene	1.187	1.139	4.0	96	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6