

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121819\
 Data File : VX014089.D
 Acq On : 18 Dec 2019 12:17
 Operator : JC/SP
 Sample : K6252-02DL 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE132D4-20191209DL

Quant Time: Dec 18 16:48:22 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 17 03:01:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	563457	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	868973	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	761597	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	327209	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	312297	48.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.80%	
35) Dibromofluoromethane	5.49	113	258364	48.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.82%	
50) Toluene-d8	8.71	98	1018390	49.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.04%	
62) 4-Bromofluorobenzene	11.13	95	334825	44.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.08%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	2.38	101	6853	1.287	ug/l	96
44) Trichloroethene	7.21	130	273032	40.293	ug/l	100
52) Toluene	8.78	92	18899	1.219	ug/l	92
64) Tetrachloroethene	9.33	164	5592	0.830	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121819\
Data File : VX014089.D
Acq On : 18 Dec 2019 12:17
Operator : JC/SP
Sample : K6252-02DL 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE132D4-20191209DL

Quant Time: Dec 18 16:48:22 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 17 03:01:07 2019
Response via : Initial Calibration

