

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121819\
 Data File : VX014091.D
 Acq On : 18 Dec 2019 13:04
 Operator : JC/SP
 Sample : K6251-07DL 50X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE105D2-20191206DL

Quant Time: Dec 18 16:56:18 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 17 03:01:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	559050	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	852460	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	770463	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	333307	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	313901	49.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.08%	
35) Dibromofluoromethane	5.49	113	255298	49.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.54%	
50) Toluene-d8	8.71	98	1023320	50.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.44%	
62) 4-Bromofluorobenzene	11.13	95	339037	45.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.94%	
Target Compounds						
44) Trichloroethene	7.21	130	159423	23.983	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121819\
Data File : VX014091.D
Acq On : 18 Dec 2019 13:04
Operator : JC/SP
Sample : K6251-07DL 50X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE105D2-20191206DL

Quant Time: Dec 18 16:56:18 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 17 03:01:07 2019
Response via : Initial Calibration

