

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121819\
 Data File : VX014094.D
 Acq On : 18 Dec 2019 14:14
 Operator : JC/SP
 Sample : K6252-04DL 50X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE132D6-20191209DL

Quant Time: Dec 18 16:42:05 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 17 03:01:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	553228	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	855111	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	763679	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	338894	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	310762	49.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.12%	
35) Dibromofluoromethane	5.49	113	256640	49.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.74%	
50) Toluene-d8	8.71	98	1012188	50.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.04%	
62) 4-Bromofluorobenzene	11.13	95	336925	45.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.08%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	2.38	101	3474	0.664	ug/l	95
44) Trichloroethene	7.21	130	188723	28.302	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121819\
Data File : VX014094.D
Acq On : 18 Dec 2019 14:14
Operator : JC/SP
Sample : K6252-04DL 50X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE132D6-20191209DL

Quant Time: Dec 18 16:42:05 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 17 03:01:07 2019
Response via : Initial Calibration

