

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121819\  
 Data File : VX014101.D  
 Acq On : 18 Dec 2019 16:57  
 Operator : JC/SP  
 Sample : K6300-05  
 Misc : 5.0mL/MSVOA X/WATER  
 ALS Vial : 19 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 SSI-BFD

Quant Time: Dec 19 03:33:24 2019  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_X\METHOD\82X121319W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Dec 17 03:01:07 2019  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc  | Units   | Dev(Min) |
|-----------------------------|-------|------|----------|-------|---------|----------|
| 1) Pentafluorobenzene       | 5.65  | 168  | 559625   | 50.00 | ug/l    | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.85  | 114  | 864762   | 50.00 | ug/l    | 0.00     |
| 63) Chlorobenzene-d5        | 10.11 | 117  | 775172   | 50.00 | ug/l    | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.07 | 152  | 329046   | 50.00 | ug/l    | 0.00     |
| System Monitoring Compounds |       |      |          |       |         |          |
| 33) 1,2-Dichloroethane-d4   | 6.05  | 65   | 316669   | 49.93 | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =     | 99.86%  |          |
| 35) Dibromofluoromethane    | 5.49  | 113  | 260941   | 49.64 | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =     | 99.28%  |          |
| 50) Toluene-d8              | 8.71  | 98   | 1029753  | 50.32 | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =     | 100.64% |          |
| 62) 4-Bromofluorobenzene    | 11.13 | 95   | 338947   | 45.31 | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =     | 90.62%  |          |
| Target Compounds            |       |      |          |       |         |          |
| 16) Acetone                 | 2.43  | 43   | 16999    | 4.128 | ug/l    | 97       |
| 32) 1,1,1-Trichloroethane   | 5.49  | 97   | 10909    | 1.267 | ug/l #  | 50       |
| 64) Tetrachloroethene       | 9.33  | 164  | 6426     | 0.937 | ug/l    | 93       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121819\  
Data File : VX014101.D  
Acq On : 18 Dec 2019 16:57  
Operator : JC/SP  
Sample : K6300-05  
Misc : 5.0mL/MSVOA X/WATER  
ALS Vial : 19 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
SSI-BFD

Quant Time: Dec 19 03:33:24 2019  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_X\METHOD\82X121319W.M  
Quant Title : SW846 8260  
QLast Update : Tue Dec 17 03:01:07 2019  
Response via : Initial Calibration

