

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121820\
 Data File : VX020124.D
 Acq On : 18 Dec 2020 14:05
 Operator : JC/MD
 Sample : VX1218WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1218WBL01

Quant Time: Dec 18 14:42:10 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 18 13:09:49 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	323954	50.00	ug/l	0.01
34) 1,4-Difluorobenzene	6.824	114	540020	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.092	117	489436	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	212927	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.025	65	202266	46.58	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	93.16%
35) Dibromofluoromethane	5.458	113	165386	47.88	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.76%
50) Toluene-d8	8.696	98	642570	48.32	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.64%
62) 4-Bromofluorobenzene	11.122	95	231733	45.53	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	91.06%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121820\
Data File : VX020124.D
Acq On : 18 Dec 2020 14:05
Operator : JC/MD
Sample : VX1218WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1218WBL01

Quant Time: Dec 18 14:42:10 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 18 13:09:49 2020
Response via : Initial Calibration

