

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX121820\
 Data File : VX020141.D
 Acq On : 19 Dec 2020 00:46
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 19 04:51:27 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 18 13:09:49 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	100	0.00
2 T	Dichlorodifluoromethane	50.000	46.559	6.9	99	0.00
3 P	Chloromethane	50.000	52.523	-5.0	106	0.00
4 C	Vinyl Chloride	50.000	47.353	5.3#	101	0.00
5 T	Bromomethane	50.000	55.258	-10.5	116	0.00
6 T	Chloroethane	50.000	52.758	-5.5	105	0.00
7 T	Trichlorofluoromethane	50.000	46.504	7.0	101	0.00
8 T	Diethyl Ether	50.000	57.470	-14.9	107	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	45.824	8.4	100	0.00
10 T	Methyl Iodide	50.000	51.367	-2.7	105	0.00
11 T	Tert butyl alcohol	250.000	279.765	-11.9	116	0.00
12 CM	1,1-Dichloroethene	50.000	47.177	5.6#	101	0.00
13 T	Acrolein	250.000	274.507	-9.8	108	0.00
14 T	Allyl chloride	50.000	52.082	-4.2	102	0.00
15 T	Acrylonitrile	250.000	290.223	-16.1	108	0.00
16 T	Acetone	250.000	254.662	-1.9	101	0.00
17 T	Carbon Disulfide	50.000	46.286	7.4	99	0.00
18 T	Methyl Acetate	50.000	57.960	-15.9	110	0.00
19 T	Methyl tert-butyl Ether	50.000	58.651	-17.3	107	0.00
20 T	Methylene Chloride	50.000	54.005	-8.0	106	0.00
21 T	trans-1,2-Dichloroethene	50.000	49.674	0.7	100	0.00
22 T	Diisopropyl ether	50.000	57.399	-14.8	106	0.00
23 T	Vinyl Acetate	250.000	298.881	-19.6	106	0.00
24 P	1,1-Dichloroethane	50.000	53.893	-7.8	105	0.00
25 T	2-Butanone	250.000	283.645	-13.5	106	0.00
26 T	2,2-Dichloropropane	50.000	43.579	12.8	88	0.00
27 T	cis-1,2-Dichloroethene	50.000	53.709	-7.4	103	0.00
28 T	Bromochloromethane	50.000	55.977	-12.0	108	0.00
29 T	Tetrahydrofuran	250.000	294.666	-17.9	109	0.00
30 C	Chloroform	50.000	54.094	-8.2#	103	0.00
31 T	Cyclohexane	50.000	46.260	7.5	99	0.00
32 T	1,1,1-Trichloroethane	50.000	49.116	1.8	100	0.00
33 S	1,2-Dichloroethane-d4	50.000	56.339	-12.7	108	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	99	0.00
35 S	Dibromofluoromethane	50.000	53.510	-7.0	106	0.00
36 T	1,1-Dichloropropene	50.000	45.346	9.3	99	0.00
37 T	Ethyl Acetate	50.000	56.159	-12.3	109	0.00
38 T	Carbon Tetrachloride	50.000	46.297	7.4	99	0.00
39 T	Methylcyclohexane	50.000	44.933	10.1	102	0.00
40 TM	Benzene	50.000	49.924	0.2	100	0.00
41 T	Methacrylonitrile	50.000	57.773	-15.5	107	0.00
42 TM	1,2-Dichloroethane	50.000	54.591	-9.2	105	0.00
43 T	Isopropyl Acetate	50.000	56.275	-12.5	107	0.00
44 TM	Trichloroethene	50.000	45.840	8.3	96	0.00
45 C	1,2-Dichloropropane	50.000	52.948	-5.9#	103	0.00
46 T	Dibromomethane	50.000	53.431	-6.9	104	0.00
47 T	Bromodichloromethane	50.000	54.053	-8.1	103	0.00
48 T	Methyl methacrylate	50.000	57.228	-14.5	107	0.00
49 T	1,4-Dioxane	1000.000	1147.629	-14.8	110	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121820\
 Data File : VX020141.D
 Acq On : 19 Dec 2020 00:46
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 19 04:51:27 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 18 13:09:49 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)

50 S	Toluene-d8	50.000	48.334	3.3	98	0.00
51 T	4-Methyl-2-Pentanone	250.000	285.166	-14.1	107	0.00
52 CM	Toluene	50.000	48.487	3.0#	96	0.00
53 T	t-1,3-Dichloropropene	50.000	52.972	-5.9	99	0.00
54 T	cis-1,3-Dichloropropene	50.000	52.530	-5.1	98	0.00
55 T	1,1,2-Trichloroethane	50.000	53.940	-7.9	104	0.00
56 T	Ethyl methacrylate	50.000	57.463	-14.9	105	0.00
57 T	1,3-Dichloropropane	50.000	53.977	-8.0	103	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	278.379	-11.4	105	0.00
59 T	2-Hexanone	250.000	282.124	-12.8	106	0.00
60 T	Dibromochloromethane	50.000	54.455	-8.9	102	0.00
61 T	1,2-Dibromoethane	50.000	54.405	-8.8	104	0.00
62 S	4-Bromofluorobenzene	50.000	47.613	4.8	96	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	95	0.00
64 T	Tetrachloroethene	50.000	45.317	9.4	96	0.00
65 PM	Chlorobenzene	50.000	48.310	3.4	94	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	51.350	-2.7	98	0.00
67 C	Ethyl Benzene	50.000	47.577	4.8#	93	0.00
68 T	m/p-Xylenes	100.000	96.128	3.9	92	0.00
69 T	o-Xylene	50.000	49.060	1.9	92	0.00
70 T	Styrene	50.000	50.243	-0.5	92	0.00
71 P	Bromoform	50.000	55.239	-10.5	104	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	90	0.00
73 T	Isopropylbenzene	50.000	49.710	0.6	93	0.00
74 T	N-aryl acetate	50.000	60.630	-21.3	103	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	57.055	-14.1	104	0.00
76 T	1,2,3-Trichloropropane	50.000	57.462	-14.9	105	0.00
77 T	Bromobenzene	50.000	49.874	0.3	94	0.00
78 T	n-propylbenzene	50.000	49.883	0.2	94	0.00
79 T	2-Chlorotoluene	50.000	49.513	1.0	91	0.00
80 T	1,3,5-Trimethylbenzene	50.000	50.630	-1.3	92	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	57.043	-14.1	99	0.00
82 T	4-Chlorotoluene	50.000	49.598	0.8	91	0.00
83 T	tert-Butylbenzene	50.000	50.231	-0.5	94	0.00
84 T	1,2,4-Trimethylbenzene	50.000	50.823	-1.6	90	0.00
85 T	sec-Butylbenzene	50.000	50.371	-0.7	97	0.00
86 T	p-Isopropyltoluene	50.000	50.955	-1.9	95	0.00
87 T	1,3-Dichlorobenzene	50.000	47.639	4.7	88	0.00
88 T	1,4-Dichlorobenzene	50.000	46.788	6.4	89	0.00
89 T	n-Butylbenzene	50.000	49.911	0.2	95	0.00
90 T	Hexachloroethane	50.000	48.944	2.1	94	0.00
91 T	1,2-Dichlorobenzene	50.000	49.351	1.3	91	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	58.754	-17.5	103	0.00
93 T	1,2,4-Trichlorobenzene	50.000	49.616	0.8	92	0.00
94 T	Hexachlorobutadiene	50.000	49.552	0.9	108	0.00
95 T	Naphthalene	50.000	56.503	-13.0	97	0.00
96 T	1,2,3-Trichlorobenzene	50.000	51.782	-3.6	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121820\
Data File : VX020141.D
Acq On : 19 Dec 2020 00:46
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Dec 19 04:51:27 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820w.M
Quant Title : SW846 8260
QLast Update : Fri Dec 18 13:09:49 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 6