

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121821\
 Data File : VX025878.D
 Acq On : 17 Dec 2021 12:10
 Operator : JC/MD
 Sample : VX1218MBL01
 Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1218MBL01

Quant Time: Dec 19 02:33:13 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 14 07:02:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	145833	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	248394	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	221056	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	90989	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	93904	44.753	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	89.500%
35) Dibromofluoromethane	5.385	113	78665	45.410	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.820%
50) Toluene-d8	8.647	98	292002	46.263	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.520%
62) 4-Bromofluorobenzene	11.085	95	98759	41.191	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	82.380%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121821\
Data File : VX025878.D
Acq On : 17 Dec 2021 12:10
Operator : JC/MD
Sample : VX1218MBL01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1218MBL01

Quant Time: Dec 19 02:33:13 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 14 07:02:11 2021
Response via : Initial Calibration

