

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121821\
 Data File : VX025882.D
 Acq On : 17 Dec 2021 13:55
 Operator : JC/MD
 Sample : M5137-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GWBR3D-G-115-140-121621

Quant Time: Dec 19 02:28:08 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 14 07:02:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	142682	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	243786	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	220574	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	93441	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	103837	50.580	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 101.160%		
35) Dibromofluoromethane	5.391	113	84466	49.680	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 99.360%		
50) Toluene-d8	8.653	98	309728	49.999	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 100.000%		
62) 4-Bromofluorobenzene	11.085	95	106181	45.123	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 90.240%		
Target Compounds						Qvalue
12) 1,1-Dichloroethene	2.319	96	13365	8.848	ug/l	87
21) trans-1,2-Dichloroethene	3.099	96	2719	1.613	ug/l	87
24) 1,1-Dichloroethane	3.617	63	5869	1.922	ug/l #	90
27) cis-1,2-Dichloroethene	4.495	96	138886	69.484	ug/l	92
42) 1,2-Dichloroethane	6.092	62	2988	1.143	ug/l	91
44) Trichloroethene	7.129	130	102202	53.375	ug/l	90
52) Toluene	8.720	92	17260	3.939	ug/l	99
64) Tetrachloroethene	9.275	164	629	0.357	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121821\
Data File : VX025882.D
Acq On : 17 Dec 2021 13:55
Operator : JC/MD
Sample : M5137-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GWBR3D-G-115-140-121621

Quant Time: Dec 19 02:28:08 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 14 07:02:11 2021
Response via : Initial Calibration

