

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121821\
 Data File : VX025884.D
 Acq On : 17 Dec 2021 14:42
 Operator : JC/MD
 Sample : M5137-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-01-121621

Quant Time: Dec 19 02:28:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 14 07:02:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	129435	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	228698	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	207029	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	86841	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	97291	52.241	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.480%
35) Dibromofluoromethane	5.391	113	77143	48.367	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.740%
50) Toluene-d8	8.653	98	288927	49.718	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.440%
62) 4-Bromofluorobenzene	11.085	95	99463	45.057	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.120%
Target Compounds						
44) Trichloroethene	7.135	130	458	0.255	ug/l	Qvalue 81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121821\
Data File : VX025884.D
Acq On : 17 Dec 2021 14:42
Operator : JC/MD
Sample : M5137-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-01-121621

Quant Time: Dec 19 02:28:50 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 14 07:02:11 2021
Response via : Initial Calibration

