

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121821\
 Data File : VX025885.D
 Acq On : 17 Dec 2021 15:06
 Operator : JC/MD
 Sample : M5137-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB-01-121621

Quant Time: Dec 19 02:29:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 14 07:02:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	127435	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	228742	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	206134	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	85046	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	98825	53.898	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 107.800%	
35) Dibromofluoromethane	5.391	113	78545	49.236	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 98.480%	
50) Toluene-d8	8.653	98	293175	50.440	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.880%	
62) 4-Bromofluorobenzene	11.085	95	100009	45.296	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 90.600%	
Target Compounds						Qvalue
20) Methylene Chloride	2.794	84	6553	1.849	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121821\
Data File : VX025885.D
Acq On : 17 Dec 2021 15:06
Operator : JC/MD
Sample : M5137-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB-01-121621

Quant Time: Dec 19 02:29:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 14 07:02:11 2021
Response via : Initial Calibration

